

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd Beccles Business Park  Anson Way Ellough Beccles Suffolk NR34 7TL CONTACT Henry Turner	MAKE & MODEL Ian Fellows Lucid  SERIAL NO. 215676 CAPACITY 50000kg DIVISION 20kg LOCATION  I.D. TOLERANCE +/- 3 divisions	CERT NO. 563552A  TESTED BY Scott Mitchell DATE TESTED July 2018 NEXT TEST DUE July 2019	SET NO. / CERT NO. Test Unit - EWT: TR10352/TR10406/ TR10421

**AS FOUND**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0kg	0kg	0kg	40000kg	40000kg	0kg
6000kg	6000kg	0kg	50000kg	50000kg	0kg
10000kg	10000kg	0kg			
16000kg	16000kg	0kg			
22000kg	22000kg	0kg			
30000kg	30000kg	0kg			

**DEFINITIVE**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0kg	0kg	0kg	40000kg	40000kg	0kg
6000kg	6000kg	0kg	50000kg	50000kg	0kg
10000kg	10000kg	0kg			
16000kg	16000kg	0kg			
22000kg	22000kg	0kg			
30000kg	30000kg	0kg			


**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	10000kg	10000kg	0kg
2	10000kg	10000kg	0kg
3	10000kg	10000kg	0kg
4	10000kg	10000kg	0kg

**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	10000kg	10000kg	0kg
2	10000kg	10000kg	0kg
3	10000kg	10000kg	0kg
4	10000kg	10000kg	0kg

**LAYOUT**



**COMMENTS OR OBSERVATIONS**

Needs clearing out. Very compacted.

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord Position. Service Manager Signature.

