

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd Beccles Business Park Anson Way Ellough Beccles Suffolk NR34 7TL CONTACT Henry Turner	MAKE & MODEL A12 Platform Scale SERIAL NO. 81027002 CAPACITY 1500kg DIVISION 0.5kg LOCATION Bagging I.D. TOLERANCE +/- 3 divisions	CERT NO. 563622J TESTED BY Harry Scott DATE TESTED July 2018 NEXT TEST DUE July 2019	SET NO. / CERT NO. STC: UH0094/UH0236 STC: UG1598/UH0381

**AS FOUND**

LOAD POINT	READING	ERROR + / -	LOAD POINT	READING	ERROR + / -
0.0kg	0.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			
500.0kg	500.0kg	0.0kg			
1000.0kg	1000.5kg	+ 0.5kg			
1500.0kg	1500.5kg	+ 0.5kg			

**DEFINITIVE**

LOAD POINT	READING	ERROR + / -	LOAD POINT	READING	ERROR + / -
0.0kg	0.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			
500.0kg	500.0kg	0.0kg			
1000.0kg	1000.0kg	0.0kg			
1500.0kg	1500.0kg	0.0kg			


**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR + / -
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR + / -
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

**LAYOUT**



COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord Position. Service Manager Signature.

