

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd Beccles Business Park Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u> Michael Toleman	<u>MAKE & MODEL</u> Lucid & Weighbridge <u>SERIAL NO.</u> 215676 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 566397A <u>TESTED BY</u> Scott Mitchell <u>DATE TESTED</u> July 2019 <u>NEXT TEST DUE</u> July 2020	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR10624/TJ0005

AS FOUND

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0kg	0kg	0kg	40000kg	40000kg	0kg
6000kg	6000kg	0kg	50000kg	50000kg	0kg
10000kg	10000kg	0kg			
16000kg	16000kg	0kg			
22000kg	22000kg	0kg			
30000kg	30000kg	0kg			


DEFINITIVE

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0kg	0kg	0kg	40000kg	40000kg	0kg
6000kg	6000kg	0kg	50000kg	50000kg	0kg
10000kg	10000kg	0kg			
16000kg	16000kg	0kg			
22000kg	22000kg	0kg			
30000kg	30000kg	0kg			

CORNERING AS FOUND

CORNER No.	LOAD POINT	READING	ERROR +/-
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

LAYOUT



CORNERING DEFINITIVE

CORNER No.	LOAD POINT	READING	ERROR +/-
1	7000kg	7000kg	0kg
2	7000kg	7000kg	0kg
3	7000kg	7000kg	0kg
4	7000kg	7000kg	0kg
5	7000kg	7000kg	0kg
6	7000kg	7000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord **Position.** Service Manager **Signature.** 