

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd Beccles Business Park Anson Way Ellough  Beccles Suffolk NR34 7TL <u>CONTACT</u>  Michael Toleman	<u>MAKE &amp; MODEL</u> A12 Platform Scale <u>SERIAL NO.</u> 81027002 <u>CAPACITY</u> 1500kg  <u>DIVISION</u> 0.5kg <u>LOCATION</u> Bagging  <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 566396A <u>TESTED BY</u> Scott Mitchell  <u>DATE TESTED</u> July 2019 <u>NEXT TEST DUE</u> July 2020	<u>SET NO. / CERT NO.</u> STC: UJ0191/UJ0347 STC: UH1899/UH0381 Test Unit - EWT: TR10624/TJ0005

**AS FOUND**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0.0kg	0.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			
500.0kg	500.0kg	0.0kg			
1000.0kg	1000.0kg	0.0kg			
1500.0kg	1500.0kg	0.0kg			


**DEFINITIVE**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0.0kg	0.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			
500.0kg	500.0kg	0.0kg			
1000.0kg	1000.0kg	0.0kg			
1500.0kg	1500.0kg	0.0kg			

**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

**LAYOUT**



**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

**Name.** Peter Ord    **Position.** Service Manager    **Signature.** 