

# **RADICAL**waste

**WEIGHBRIDGE & SCALES CALLIBRATION CERTIFICATES**

**WEIGHBRIDGE IN CERTIFICATE**

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park  Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u> Julian Knights	<u>MAKE &amp; MODEL</u> Lucid & Weighbridge  <u>SERIAL NO.</u> 103719 <u>CAPACITY</u> 50000kg <u>DIVISION</u> 20kg <u>LOCATION</u> Inbound  <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 572026H  <u>TESTED BY</u> Harry Scott <u>DATE TESTED</u> June 2021 <u>NEXT TEST DUE</u> June 2022	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR11160/TK0057

**AS FOUND**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0kg	0kg	0kg	30000kg	30000kg	0kg
4000kg	4000kg	0kg	40000kg	40000kg	0kg
8000kg	8000kg	0kg	50000kg	50000kg	0kg
12000kg	12000kg	0kg			
16000kg	16000kg	0kg			
20000kg	20000kg	0kg			


**DEFINITIVE**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0kg	0kg	0kg	30000kg	30000kg	0kg
4000kg	4000kg	0kg	40000kg	40000kg	0kg
8000kg	8000kg	0kg	50000kg	50000kg	0kg
12000kg	12000kg	0kg			
16000kg	16000kg	0kg			
20000kg	20000kg	0kg			

**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

**LAYOUT**




**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	16000kg	16000kg	0kg
2	16000kg	16000kg	0kg
3	16000kg	16000kg	0kg
4	16000kg	16000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

**Name.** Peter Ord    **Position.** Service Manager    **Signature.** 

**WEIGHBRIDGE OUT CERTIFICATE**

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park  Anson Way Ellough Beccles Suffolk NR34 7TL CONTACT Julian Knights	MAKE & MODEL Lucid & Weighbridge  SERIAL NO. 215676 CAPACITY 50000kg DIVISION 20kg LOCATION Outbound  I.D. TOLERANCE +/- 3 divisions	CERT NO. 572024H  TESTED BY Harry Scott DATE TESTED June 2021 NEXT TEST DUE July 2022	SET NO. / CERT NO. Test Unit - EWT: TR11160/TK0057

**AS FOUND**

LOAD POINT	READING	ERROR + / -	LOAD POINT	READING	ERROR + / -
VERIFICATION CARRIED OUT					


**DEFINITIVE**

LOAD POINT	READING	ERROR + / -	LOAD POINT	READING	ERROR + / -
0kg	0kg	0kg	40000kg	40000kg	0kg
6000kg	6000kg	0kg	50000kg	50000kg	0kg
10000kg	10000kg	0kg			
16000kg	16000kg	0kg			
22000kg	22000kg	0kg			
30000kg	30000kg	0kg			

**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR + / -
NO AS FOUND			

**LAYOUT**



**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR + / -
1	10000kg	10000kg	0kg
2	10000kg	10000kg	0kg
3	10000kg	10000kg	0kg
4	10000kg	10000kg	0kg

COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

Name. Peter Ord Position. Service Manager Signature.



## **SCALES CERTIFICATE**

CUSTOMER	MACHINE DETAILS	CERTIFICATE	TEST WEIGHTS
B & B Skip Hire Ltd. Beccles Business Park  Anson Way Ellough Beccles Suffolk NR34 7TL <u>CONTACT</u> Julian Knights	<u>MAKE &amp; MODEL</u> A12 Platform Scale  <u>SERIAL NO.</u> 81027002 <u>CAPACITY</u> 1500kg <u>DIVISION</u> 0.5kg <u>LOCATION</u> Bagging  <u>I.D</u> <u>TOLERANCE</u> +/- 3 divisions	<u>CERT NO.</u> 572025H  <u>TESTED BY</u> Harry Scott <u>DATE TESTED</u> June 2021 <u>NEXT TEST DUE</u> July 2022	<u>SET NO. / CERT NO.</u> Test Unit - EWT: TR11160/TK0057

**AS FOUND**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0.0kg	0.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			
500.0kg	500.0kg	0.0kg			
1000.0kg	1000.0kg	0.0kg			
1500.0kg	1500.0kg	0.0kg			

**DEFINITIVE**

LOAD POINT	READING	ERROR +/-	LOAD POINT	READING	ERROR +/-
0.0kg	0.0kg	0.0kg			
100.0kg	100.0kg	0.0kg			
200.0kg	200.0kg	0.0kg			
500.0kg	500.0kg	0.0kg			
1000.0kg	1000.0kg	0.0kg			
1500.0kg	1500.0kg	0.0kg			


**CORNERING AS FOUND**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

**CORNERING DEFINITIVE**

CORNER No.	LOAD POINT	READING	ERROR +/-
1	200.0kg	200.0kg	0.0kg
2	200.0kg	200.0kg	0.0kg
3	200.0kg	200.0kg	0.0kg
4	200.0kg	200.0kg	0.0kg

**LAYOUT**



COMMENTS OR OBSERVATIONS

The above test has been conducted in accordance with STC Solutions ISO 9001 certificated processes using weights traceable to National Standards.

**Name.** Peter Ord    **Position.** Service Manager    **Signature.**

